

JAN 06 2010

**VERIFICATION OF CALIBRATION REPORT**  
*of DataMaster cdm Breath Test Instrument*  
*State of Alaska*  
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:	
Name <u>J. Heinbaugh</u>	DataMaster cdm S/N <u>130173</u> ✓
ID# <u>5828</u>	Date <u>1-5-10</u>
A Agency <u>AST</u> Phone # <u>820-3263</u>	
Instrument Location <u>Glennallen</u>	
B Alco S/N <u>X124669</u> Target Value <u>0.083</u> High Pressure <u>1200</u>	
Alco Test Values	1 <sup>st</sup> Alco <u>0.082</u> ✓
	2 <sup>nd</sup> Alco <u>0.082</u> ✓
Signature <u>[Signature]</u> ✓	
COB 4/10/08 1/13/10	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of January, 2010.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130173 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173 ✓

JANUARY 05, 2010

OPERATOR'S NAME:

HEINERBAUGH

OPERATOR'S NUMBER: 5828

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: UOC

DEPT/AGENCY: 6NKG

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: 0.083 ✓

ALCO S/N: X124669

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 28.74 in ✓  
ALCO TARGET .079 10:56  
BLANK TEST .000 10:58  
INTERNAL STANDARD VERIFIED 10:58  
ALCO TU 28.73 in .082 10:58  
BLANK TEST .000 10:59  
SUBJECT SAMPLE .000 10:59  
BLANK TEST .000 11:00  
ALCO TU 28.74 in .082 11:00  
BLANK TEST .000 11:01

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173 ✓

JANUARY 05, 2010 ✓

TIME 11:01

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 28.74 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; ( = ) ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ■